



α -Amylase Assay Kit

Catalog Number KA6685

100 assays

Version: 01

Intended for research use only

www.abnova.com

Table of Contents

Introduction	3
Intended Use	3
Background	3
Principle of the Assay	3
General Information	4
Materials Supplied	4
Storage Instruction	4
Materials Required but Not Supplied	4
Precautions for Use	4
Assay Protocol	5
Assay Procedure	5
Data Analysis.....	6
Calculation of Results.....	6
Resources.....	7
References	7

Introduction

Intended Use

- Application
- ✓ Determination of α -amylase activity in blood, saliva, urine, grains and other agricultural samples.
- Features:
- ✓ Sensitive and accurate: Linear detection range 0.3 to 50 U/L α amylase in 96-well plate assay.
- ✓ Convenient: The procedure involves adding a single working reagent, incubation for 15 min, followed by the detection reagent and a 20-min incubation and reading the optical density at 585 nm.

Background

AMYLASE belongs to the family of glycoside hydrolase enzymes that break down starch into glucose molecules by acting on α -1,4- glycosidic bonds. The α -amylases (EC 3.2.1.1) cleave at random locations on the starch chain, ultimately yielding maltotriose and maltose, glucose and "limit dextrin" from amylose and amylopectin. In mammals, α -amylase is a major digestive enzyme. Increased enzyme levels in humans are associated with salivary trauma, mumps due to inflammation of the salivary glands, pancreatitis and renal failure.

Principle of the Assay

Simple, direct and automation-ready procedures for measuring amylase activity are very desirable. The α -amylase assay method involves two steps: (1). α -amylase in the sample hydrolyzes starch and the product is rapidly converted to glucose by α -glucosidase and hydrogen peroxide by glucose oxidase; (2). hydrogen peroxide concentration is determined with a colorimetric reagent.

General Information

Materials Supplied

List of component

Component	Amount
Assay Buffer (pH 7.0)	20 mL
Detection Reagent	20 mL
Glucose Standard	1 mL
Substrate	120 μ L
Enzyme A	120 μ L
Enzyme B	120 μ L

Storage Instruction

Kit is shipped on ice. Store all components at -20°C . Shelf life: 6 months after receipt.

Materials Required but Not Supplied

Pipetting devices, centrifuge tubes, clear flat-bottom 96-well plates, plate reader, and optionally membrane filters (e.g. Microcon YM-10 from Millipore).

Precautions for Use

- ✓ Reagents are for research use only. Normal precautions for laboratory reagents should be exercised while using the reagents. Please refer to Material Safety Data Sheet for detailed information.

Assay Protocol

Assay Procedure

- ✓ Reagents: Equilibrate all components to room temperature. Keep thawed Enzyme Mix in a refrigerator or on ice. The substrate may have precipitates. Prior to use, vortex tube to dissolve precipitates; gentle swirl the Detection Reagent bottle.
- ✓ Sample preparation: Ideally samples are assayed fresh. When stored frozen, α -amylase is stable for one month. Ascorbic acid, heparin, EDTA, EGTA, citrate, SDS, Tris (> 8mM) and ethanol (>0.4%) interfere and should be avoided in sample preparation. If glucose is present in the sample, treat the samples as described in GENERAL CONSIDERATIONS. It is prudent to perform a pilot test with samples at various dilutions. Recommended dilution: serum 50-fold, saliva 2,000-fold in Assay Buffer prior to assay.

1. Prepare 400 μ M Glucose Standard by mixing 10 μ L of the provided (300 mg/dL) standard with 406 μ L Assay Buffer. Transfer 10 μ L Assay Buffer, 10 μ L 400 μ M glucose, and 10 μ L of each sample into separate wells of a clear flat-bottom 96-well plate.

2. Prepare enough Working Reagent for each well by mixing 40 μ L Assay Buffer, 0.5 μ L Substrate, 1 μ L Enzyme A, 1 μ L Enzyme B. Transfer 40 μ L Working Reagent to each well. Incubate for 15 min at room temperature (25°C).

3. Add 150 μ L Detection Reagent to each well. Mix and incubate for 20 min at room temperature (25°C). Read OD585nm (540-610nm) on a plate reader.

- ✓ General Considerations

For samples known to contain glucose, use a membrane filter (e.g. Microcon YM-10 from Millipore) to remove glucose: load 50 μ L sample in a Microcon YM-10 (10 kDa cutoff) and add 500 μ L Assay Buffer. Centrifuge at 14000 rpm for 30 min, check level of sample, ideally the sample level will be less than 50 μ L. Add 500 μ L Assay Buffer and repeat the centrifugation. Measure final sample volume with a pipetman and calculate dilution factor $n = \text{final sample volume}/50 \mu\text{L}$.

Data Analysis


Calculation of Results

The Amylase activity is calculated as

$$\text{Activity} = \frac{\text{OD}_{\text{SAMPLE}} - \text{OD}_{\text{BUFFER}}}{\text{OD}_{\text{STD}} - \text{OD}_{\text{BUFFER}}} \times \frac{400}{t \text{ (min)}} \times n \text{ (U/L)}$$

$\text{OD}_{\text{SAMPLE}}$, OD_{STD} and $\text{OD}_{\text{BUFFER}}$ are optical density values of the sample, the 400 μM glucose standard and Assay Buffer. t is the incubation time. $t = 15$ min in the standard protocol. n is the dilution factor ($n = 50$ for serum, 2000 for saliva). One unit of enzyme catalyzes the production of 1 μmole of glucose per min under the assay conditions.

Note: if the calculated activity is higher than 50 U/L, dilute sample in Assay Buffer and repeat assay. Multiply the results by the dilution factor.



Standard Curve in 96-well plate assay

Resources

References

1. Han, M. J. (2020). Novel bacterial surface display system based on the Escherichia coli protein mipa. *Journal of Microbiology and Biotechnology*. 30(7): 1097-1103.
2. Bae, G.-S. (2020). Protective effect of nypa fruticans wurmb. Water extract on acute pancreatitis. *Journal of Physiology & Pathology in Korean Medicine*. 34(6): 334-340.
3. Lee, Sang-Bum, et al (2019). Impacts of whey protein on starch digestion in rumen and small intestine of steers. *Journal of Animal Science and Technology* 61.2: 98-108.